

Date: 05.03.2025

Page: 1 of 1

**Inspection certificate 3.1
according to DIN EN 10204**



Powering Business Worldwide

We herewith confirm that the delivered goods correspond to high quality requirements and comply with below mentioned characteristics.

Customer: BHF Technologies Pty Ltd
4 Kingston Town Close
Oakleigh
Victoria 3166

Product: 29450491000079 BECO PR Endura S

Size: 491 514 x 633 mm, 6 LongH

Lot-No.: 18052

Ref: PO004811

Production date: 13.02.2025

Expiry date: 13.02.2028

Specifications:

Test criterion	Unit	Set value	Tolerance	Measured Value *
Thickness	[mm]	1,50	± 0,20	1,62
Water flow (1,0 bar)	[L/m ² min]	770	± 120	783
Bursting strenght, wet	[kPa]	> 400		790
Endotoxin content	[EU/mL]	< 0,08		< 0,08

Test criterion	Unit	Target value
Ash content	[%]	0,3

* measured values are averages of the batch

Indicated test data are results of quality verifications and do not substitute buyer's quality verification upon receiving. The indicated quality characteristics do not necessarily guarantee the qualification for a particular application.

Konstantina Michalatou
Assistant quality manager
Quality Assurance

Henrik Beisiegel
Laboratory assistant manager
Quality Assurance