

„Integrity test“ of BECODISC Modules

Article number	Time to empty the housing	Test pressure set value	Stabilization time	Test time	max. pressure drop "delta p"
	[min]	[mbar]	[s]	[s]	[mbar/min]
B08S...	3	100	180	180	20
B06S...	3	100	180	180	20
B05S...	3	150	180	180	20
B04S...	3	150	180	180	20
B03S...	3	200	180	180	20
B02S...	3	200	180	180	20
B01S...	3	200	180	180	20